


 Understandings of principle, structure and influence factors WAFANGJIAN 10010101 OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIST)	 Weighing instrument Measuring instrument that serves to determine the mass of a body by using the action of gravity on this body. The instrument may also be used to determine other quantities, magnitudes, parameters or characteristic related to mass. According to its method of operation, a weighing instrument is classified as an automatic or non-automatic instrument. WAFANGJIAN 10010101 OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIST)
 Automatic weighing instrument An instrument that weighs without the intervention of an operator and follows a predetermined program of automatic processes characteristic of the instrument.  WAFANGJIAN 10010101 OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIST)	 Non-automatic weighing instrument Instrument that requires the intervention of an operator during the weighing process, for example to deposit on or remove from the receptor the load to be measured and also to obtain the result.  WAFANGJIAN 10010101 OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIST)

What's Mechanical Weighing Scale ?

Mechanical weighing scales use mechanical principle such as spring and lever when weighing is measured.

Various Mechanical Weighing Scale



Measurement principle

Lever

Elasticity


Spring

Load cell (electronically)

Levers in Mechanical Weighing Instruments

First order levers

The fulcrum is between the Load and Resistant. The resistant is usually either to the load or is less. In the case the lever has a mechanical advantage.



Second order levers

The fulcrum is at one end of the lever and the resistant is at the other. The lever has a mechanical advantage.




Third order levers


The fulcrum is at one end of the lever and the resistant is between the fulcrum and load. The lever has a mechanical advantage.




Combination of levers



Spring type



Elasticity : Spring



Hooke's law

Growth of the spring is in proportion to increased force

$$F = -kx$$

$$k = \frac{P}{\delta} = \frac{Gd^4}{8nd^3}$$

F : Restoring force exerted by the spring

k : Spring constant (N/mm)

x : The distance by which the spring is elongated

P : Load (N)

δ : Displacement (mm)

G : Rigidity of spring materials (N/mm²)

d : Diameter of spring (mm)

n : Total number

D : Diameter of the coil


Spring -Creep and Hysteric's-

Creep



The strain by the outside power changes with the passage of time when the outside power is left while added.

Hystericis



The difference is caused in the relation between the stress and the strain in the increase and decrease of the stress.

Spring -Temperature compensating-


The elastic coefficient changes with the temperature. The strain to constant outside power is different depending on the temperature.

Temperature compensating device


Bimetal

Compensated spring

Equal arm balance - Sensitivity -




Equal arm balance - Eccentricity -



Roberval mechanism


Simple lever



Same mass and same ratio of "ab" and "bc" $\Rightarrow$ **Equilibrium**


Same mass and difference ratio of "ab" and "bc" $\Rightarrow$ **Not Equilibrium**

Roberval mechanism




It's form two parallelogram "abde", "bcdf" which connected the point of length and breadth with a pin, and move freely.
 "abc" becomes level when it's put the same mass on a plate in the extension of "ad", "ef".
 "abc" becomes always level even if it's move a position of the mass then on a plate.

Use example of the Roberval mechanism



Robarval Eccentricity



Example:

$a = 60 \text{ mm}$

$d = 180 \text{ mm}$

$\sigma = 1 \text{ mm}$


$W = 2.5 \text{ kg}$

$e = 50 \text{ mm}$

$E = ?$

$$E = \frac{We\sigma}{ad}$$

Robarval Eccentricity



Answer :

$$E = \frac{2500 \times 50 \times 1}{60 \times 180} = 11.6 \text{ g}$$

$$E = \frac{We\sigma}{ad}$$

Gravity compensation

All objects on the earth are subject to Earth's gravity all the time.
No measurements of force and mass can be free from the gravity effects, as the force and mass are measured by balancing the object's weight on a lever, or by converting it into modification of elastic body.

Gravity compensation is required because the gravity depends on the location on the earth (ex. latitude, longitude, altitude). 9.80665m/s² is used as the standard gravitational acceleration.

If a measuring instrument equipped with elastic bodies is calibrated at a place different from a place of use, the value of gravity compensation E_g is obtained in the following equation:

$$E_g = \frac{g_1 - g_2}{9.8} \cdot W$$

Where:

g_1 : gravitational acceleration at the calibration site

g_2 : gravitational acceleration at the place of use

W : mass, corresponding to inspection weight

Gravity compensation

A range of the gravity in Japan

9.806 m/s² ~ 9.791 m/s²
Hokkaido: 9.806 m/s² ... (g_2)



Tsukuba 9.800 m/s² ... (g_1)

EX)

$$I_1 \propto W \cdot (g_1 = 9.800 \text{ m/s}^2)$$

$$I_2 \propto W \cdot (g_2 = 9.806 \text{ m/s}^2)$$

$$I_1/I_2 = g_1/g_2$$

$$I_1 = I_2 \cdot g_1/g_2$$


$$= 5.000 \cdot 9.806/9.800 = 5.003, I_1 = 5.003 \text{ (kg)}$$

Okinawa 9.791 m/s²

 Verification procedure for NAWIs 【 OIML R76 Edition 1992 】 INTERNATIONAL BUREAU OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (IAIST)	 Contents of OIML R76 1992 INTERNATIONAL BUREAU OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (IAIST)
 The related term of the mechanical weighing scale INTERNATIONAL BUREAU OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (IAIST)	 T.1.2 Non automatic weighing instrument INTERNATIONAL BUREAU OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (IAIST)

T.1.2.1 Graduated instrument

Instrument allowing the direct reading of the complete or partial weighing result.




T.1.2.2 Non graduated instrument

Instrument not fitted with a scale numbered in units of mass.



T.1.2.3 Self indicating instrument


Instrument in which the position of equilibrium is obtained without the intervention of an operator.



T.1.2.4 Semi self indicating instrument

Instrument with a self indication weighing range, in which the operator intervenes to alter the limits of this range.

Example




Measurement procedure

1. Load applies to load receptor
2. Remove inside weights by a knob
3. Repeat 2 till it is balanced in a weighing range
4. Load =
Remove value (knob) + reading for indicator
Knob = 40g
Indicator = 4.3g
Load = 44.3g

T.1.2.5 Non self indicating instrument


Instrument in which the position of equilibrium is obtained entirely by operator.

Example



Measurement procedure

1. Load applies to load receptor
2. Weight applies to take the rough equilibrium
3. Sliding moves to complete equilibrium
4. Load =
total weight value + reading of the sliding

 T.2.1 Main device T.2.1.1 Load receptor Part of the instrument intended to receive the load. T.2.1.2 Load-transmitting device Part of the instrument for transmitting the force produced by the load acting on the load receptor, to the load measuring device. T.2.1.3 Load measuring device Part of the instrument for measuring the mass of the load by means of an equilibrium device for balancing the force coming from the load transmitting device, and an indicating or printing device. T.2.4 Indicating device (of a weighing instrument) Part of the load measuring device on which the direct reading of the result is obtained. INTERNATIONAL BUREAU OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (IAIST)	 T.2.4.1 Indicating component Component indicating the equilibrium and/or the result. On an instrument with several positions of equilibrium it indicates only the equilibrium (so-called zero). On an instrument with several positions of equilibrium it indicates both the equilibrium and the result. On an electronic instrument, this is the display.  T.2.4.2 Scale mark A line or other mark on an indicating component corresponding to a specified value of mass. T.2.4.3 Scale base An imaginary line through the centers of all the shortest scale marks. INTERNATIONAL BUREAU OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (IAIST)
 T.2.7.2 Zero setting device Device for setting the indication to zero when there is no load on the load receptor. T.2.7.2.1 Non automatic zero setting device Device for setting the indication to zero by an operator. INTERNATIONAL BUREAU OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (IAIST)	 T.2.7.4 Tare device Device for setting the indication to zero when a load is on the load receptor: Without altering the weighing range for net loads (additive tare device), or Reducing the weighing range for net loads (subtractive tare device). It may function as: - A non automatic device (load balanced by an operator), - A semi automatic device (load balanced automatically following a single manual command), - An automatic device (load balanced automatically without the intervention of an operator). INTERNATIONAL BUREAU OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (IAIST)

 T.3.2 Scale divisions T.3.2.1 Scale spacing (<i>instrument with analogue indication</i>) Distance between any two consecutive scale marks, measured along the scale base. T.3.2.2 Actual scale interval (d) Value expressed in units of mass of: The difference between the values corresponding to two consecutive scale marks, for analogue indication, or The difference between two consecutive indication values, for digital indication. T.3.2.3 Verification scale interval (e) Value, expressed in units of mass, used for the classification and verification of an instrument T.3.2.4 Scale interval of numbering Value of the difference between two consecutive numbered scale marks. INTERNAL DOCUMENT OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIIST)	 T.3.2.5 Number of verification scale intervals (<i>single-interval instrument</i>) Quotient of the maximum capacity and the verification scale interval: $n = \text{Max} / e$ T.4 Metrological properties of an instrument T.4.1 Sensitivity (<i>k</i>) For given value of the measured mass, the quotient of the change of the observed variable <i>I</i> and the corresponding change of the measured <i>M</i> $k = \Delta I / \Delta M$ INTERNAL DOCUMENT OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIIST)
 Verification procedure for NAWIs INTERNAL DOCUMENT OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIIST)	 Verification Item (8.3 initial verification) ☐ Visual inspection (8.3.2) ☐ Tests (8.3.3) ◇ Error of indication [Weighing test] ◇ Value of maximum permissible error ◇ Maximum permissible errors for net values ◇ Tare weighing device - Accuracy of zero setting and tare devices - Repeatability - Eccentricity - Discrimination - Sensitivity of non self indication instruments INTERNAL DOCUMENT OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIIST)

 Visual inspection (8.3.2) <i>Before testing, the instrument shall be visual inspected</i> • <i>Confirmation of markings: Accuracy class, Max, Min, e, d</i> • <i>Confirmation of condition of instrument</i> • <i>Confirmation of level of instrument</i> <i>*Results of the visual inspection should be recorded in the test report, including test date and observer, etc.</i> INTERNAL DOCUMENT OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIIST)	 Tests (8.3.3) INTERNAL DOCUMENT OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIIST)
 Value of maximum permissible error Maximum permissible error for verification INTERNAL DOCUMENT OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIIST)	 Accuracy classes for NAWI's (3.2) Class 1 Special accuracy ① ultra micro-, micro-, semi micro, macro- Class 2 High accuracy ② precision balances Class 3 Medium accuracy ③ NAWI's for trade use Class 4 Ordinary accuracy ④ NAWI's for lower accuracy INTERNAL DOCUMENT OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIIST)

Accuracy classes for NAWI's (3.2)

Accuracy class	Verification scale interval	Number of verification scale intervals Minimum	Maximum	Minimum capacity
Special (I)	$0.001g \leq e$	50 000	-	100 e
High (II)	$0.001g \leq e \leq 0.05g$ $0.1g \leq e$	100 5 000	100 000 100 000	20 e 50 e
Medium (III)	$0.1g \leq e \leq 2g$ $5g \leq e$	100 500	10 000 10 000	20 e 20 e
Ordinary (IV)	$5g \leq e$	100	1 000	10 e

Verification scale interval (e)

Value, expressed in unit of mass, used for the classification and verification of an instrument.

Number of verification scale interval (n)

Quotient of the maximum capacity and the verification scale interval.

$$n = \text{Max} / e$$

Maximum permissible error (3.5.1)

Verification

Maximum permissible errors	For loads, m expressed in verification scale intervals, e			
	Class 1	Class 2	Class 3	Class 4
$\pm 0.5e$	$0 \leq m \leq 50\,000$	$0 \leq m \leq 5\,000$	$0 \leq m \leq 500$	$0 \leq m \leq 50$
$\pm 1.0e$	$50\,000 < m \leq 200\,000$	$5\,000 < m \leq 20\,000$	$500 < m \leq 2\,000$	$50 < m \leq 200$
$\pm 1.5e$	$200\,000 \leq m$	$20\,000 < m \leq 100\,000$	$2\,000 < m \leq 10\,000$	$200 < m \leq 1\,000$

Exercise

Determine MPE and MPE change points.

Exercise 1

Determine MPE and MPE change points for verification

Specification of NAWI is

**Max : 3 kg,
Verification scale Interval (e): 1 g,
Accuracy class 3**

For loads, m expressed in verification scale intervals, e


Minimum permissible errors	Class 1	Class 2	Class 3	Class 4
$\pm 0.5 e$	$0 \leq m \leq 50\,000$	$0 \leq m \leq 5\,000$	$0 \leq m \leq 500$	$0 \leq m \leq 50$
$\pm 1.0 e$	$50\,000 < m \leq 200\,000$	$5\,000 < m \leq 20\,000$	$500 < m \leq 2\,000$	$50 < m \leq 200$
$\pm 1.5 e$	$200\,000 < m$	$20\,000 < m \leq 100\,000$	$2\,000 < m \leq 10\,000$	$200 < m \leq 1\,000$

MPE (e)	Class 3 For loads, m expressed in verification scale intervals, e	MPE (g)	Class 3 (mass)
$\pm 0.5 e$	$0 \leq m \leq 500$	$\pm 0.5 \times 1g = \pm 0.5 g$	$0 \leq m \leq 500 \times 1 g (e)$ $0 g \leq m \leq 500 g$
$\pm 1.0 e$	$500 < m \leq 2\,000$	$\pm 1.0 \times 1g = \pm 1 g$	$500 \times 1g (e) < m \leq 2\,000 \times 1 g (e)$ $500 g < m \leq 2\,000 g$
$\pm 1.5 e$	$2\,000 < m \leq 10\,000$	$\pm 1.5 \times 1g = \pm 1.5 g$	$2\,000 \times 1g (e) < m \leq 10\,000 \times 1 g (e)$ $2\,000 g < m \leq 10\,000 g$

Answer

Max : 3 kg, Verification scale Interval (e): 1 g, Accuracy class 3

MPE (g)	Class 3 (mass)	Mpe change points
$\pm 0.5 g$	$0 g \leq m \leq 500 g$	500 g (change point 1)
$\pm 1.0 g$	$500 g < m \leq 2 kg$	2 kg (change point 2)
$\pm 1.5 g$	$2 kg < m \leq 3 kg (Max)$	-



Exercise 2

Determine MPE and MPE change points for verification

Specification of NAWI is

**Max : 6 kg,
Verification scale Interval (e): 2 g,
Accuracy class 3**

For loads, m expressed in verification scale intervals, e

Maximum permissible errors	Class 1	Class 2	Class 3	Class 4
$\pm 0.5 e$	$0 \leq m \leq 50\,000$	$0 \leq m \leq 5\,000$	$0 \leq m \leq 500$	$0 \leq m \leq 50$
$\pm 1.0 e$	$50\,000 < m \leq 200\,000$	$5\,000 < m \leq 20\,000$	$500 < m \leq 2\,000$	$50 < m \leq 200$
$\pm 1.5 e$	$200\,000 < m$	$20\,000 < m \leq 100\,000$	$2\,000 < m \leq 10\,000$	$200 < m \leq 1\,000$


MPE (e)	Class 3 For loads, m expressed in verification scale intervals, e	MPE (g)	Class 3 (mass)
$\pm 0.5 e$	$0 \leq m \leq 500$	$\pm 0.5 \times 2g = \pm 1.0 g$	$0 \leq m \leq 500 \times 2 g (e)$
$\pm 1.0 e$	$500 < m \leq 2\,000$	$\pm 1.0 \times 2g = \pm 2.0 g$	$500 \times 2g (e) < m \leq 2\,000 \times 2 g (e)$
$\pm 1.5 e$	$2\,000 < m \leq 10\,000$	$\pm 1.5 \times 2g = \pm 3.0 g$	$2\,000 \times 2g (e) < m \leq 10\,000 \times 2g (e)$

INTERNAL MEMO OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIST)

Answer

Max : 6 kg, Verification scale Interval (e): 2g, Accuracy class 3

MPE (g)	Class 3 (mass)	Mpe change points
$\pm 1.0 g$	$0 g \leq m \leq 1 kg$	1 kg (change point 1)
$\pm 2.0 g$	$1 kg < m \leq 4 kg$	4 kg (change point 2)
$\pm 3.0 g$	$4 kg < m \leq 6 kg (Max)$	-



INTERNAL MEMO OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIST)

Exercise 3

Determine MPE and MPE change points for verification

Specification of NAWI is


Max : 5 kg,
Verification scale Interval (e): 10 g,
Accuracy class 3 or class 4

INTERNAL MEMO OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIST)

Accuracy classes for NAWI's (3.2)

Accuracy class	Verification scale interval	Number of verification scale interval	Minimum capacity
Special (I)	$0.001 g \leq e$	Minimum 50 000	100 e
High (II)	$0.001 g \leq e \leq 0.05 g$	100	100 000
Medium (III)	$0.1 g \leq e \leq 2 g$	5 000	100 000
Ordinary (IV)	$5 g \leq e$	100	10 000

INTERNAL MEMO OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIST)



VERNAL INSTITUTE OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIST)

AIST


Case of class 3

For loads, m expressed in verification scale intervals, e

Minimum permissible errors	Class 1	Class 2	Class 3	Class 4
$\pm 0.5 e$	$0 \leq m \leq 50\,000$	$0 \leq m \leq 5\,000$	$0 \leq m \leq 500$	$0 \leq m \leq 50$
$\pm 1.0 e$	$50\,000 < m \leq 200\,000$	$5\,000 < m \leq 20\,000$	$500 < m \leq 2\,000$	$50 < m \leq 200$
$\pm 1.5 e$	$200\,000 < m$	$20\,000 < m \leq 100\,000$	$2\,000 < m \leq 10\,000$	$200 < m \leq 1\,000$

MPE (e)	Class 3 For loads, m expressed in verification scale intervals, e	MPE (e)	Class 3 (mass)
$\pm 0.5 e$	$0 \leq m \leq 500$	$\pm 0.5 \times 10 g = \pm 5 g$	$0 \leq m \leq 500 \times 10 g (e)$ $0 g \leq m \leq 5 kg$
$\pm 1.0 e$	$500 < m \leq 2\,000$	$\pm 1.0 \times 10 g = \pm 10 g$	$500 \times 10 g (e) < m \leq 2\,000 \times 10 g (e)$ $5 kg < m \leq 20 kg$
$\pm 1.5 e$	$2\,000 < m \leq 10\,000$	$\pm 1.5 \times 10 g = \pm 15 g$	$2\,000 \times 10 g (e) < m \leq 10\,000 \times 10 g (e)$ $20 kg < m \leq 100 kg$

VERNAL INSTITUTE OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIST)



AIST

Case of class 4

For loads, m expressed in verification scale intervals, e

Minimum permissible errors	Class 1	Class 2	Class 3	Class 4
$\pm 0.5 e$	$0 \leq m \leq 50\,000$	$0 \leq m \leq 5\,000$	$0 \leq m \leq 500$	$0 \leq m \leq 50$
$\pm 1.0 e$	$50\,000 < m \leq 200\,000$	$5\,000 < m \leq 20\,000$	$500 < m \leq 2\,000$	$50 < m \leq 200$
$\pm 1.5 e$	$200\,000 < m$	$20\,000 < m \leq 100\,000$	$2\,000 < m \leq 10\,000$	$200 < m \leq 1\,000$


MPE (e)	Class 4 For loads, m expressed in verification scale intervals, e	MPE (e)	Class 3 (mass)
$\pm 0.5 e$	$0 \leq m \leq 50$	$\pm 0.5 \times 10 g = \pm 5 g$	$0 \leq m \leq 50 \times 10 g (e)$ $0 g \leq m \leq 500 g$
$\pm 1.0 e$	$50 < m \leq 200$	$\pm 1.0 \times 10 g = \pm 10 g$	$50 \times 10 g (e) < m \leq 200 \times 10 g (e)$ $500 g < m \leq 2 kg$
$\pm 1.5 e$	$200 < m \leq 1\,000$	$\pm 1.5 \times 10 g = \pm 15 g$	$200 \times 10 g (e) < m \leq 1\,000 \times 10 g (e)$ $2 kg < m \leq 10 kg$

VERNAL INSTITUTE OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIST)

Answer: case of class 4

Max : 5 kg, Verification scale Interval (e): 10 g, Accuracy class 4

MPE (g)	Class 4 (mass)	Mpe change points
± 5 g	$0 \text{ g} \leq m \leq 500 \text{ g}$	500 g
± 10 g	$500 \text{ g} < m \leq 2 \text{ kg}$	2 kg
± 15 g	$2 \text{ kg} < m \leq 5 \text{ kg (Max)}$	-



Exercise 4

Determine MPE and MPE change points for verification

Specification of NAWI is

Max : 6 kg,

e_1 : 1 g (0 – 3 kg)

e_2 : 2 g (3 – 6 kg)

Class 3

* Multi interval instrument

Max : 6 kg (e_1 : 0 – 3kg, e_2 : 3 – 6 kg)

$n_1 = \text{Max}_1 / e_1 = 3 \text{ kg} / 1 \text{ g} = 3\,000$

$n_2 = \text{Max} / e_2 = 6 \text{ kg} / 2 \text{ g} = 3\,000$

Accuracy class	Verification scale interval	Number of verification scale interval	Minimum capacity
Special Class 1	$0.001 \text{ g} \leq e$	50 000	100 e
High Class 2	$0.001 \text{ g} \leq e \leq 0.05 \text{ g}$	100	20 e
Medium Class 3	$0.1 \text{ g} \leq e$	5 000	50 e
Ordinary Class 4	$5 \text{ g} \leq e$	100	20 e
	$5 \text{ g} \leq e$	1 000	10 e


Applied e_1 (1 g)

MPE (g)	Class 3 For loads, m expressed in verification scale intervals, e	MPE (g)	Class 3 (mass)
$\pm 0.5 e_1$	$0 \leq m \leq 500$	$\pm 0.5 \times 1 \text{ g} = \pm 0.5 \text{ g}$	$0 \leq m \leq 500 \times 1 \text{ g} (e_1)$ $0 \text{ g} \leq m \leq 500 \text{ g}$
$\pm 1.0 e_1$	$500 < m \leq 2\,000$	$\pm 1.0 \times 1 \text{ g} = \pm 1 \text{ g}$	$500 \times 1 \text{ g} (e_1) < m \leq 2\,000 \times 1 \text{ g} (e_1)$ $500 \text{ g} < m \leq 2\,000 \text{ g}$
$\pm 1.5 e_1$	$2\,000 < m < 10\,000$	$\pm 1.5 \times 1 \text{ g} = \pm 1.5 \text{ g}$	$2\,000 \times 1 \text{ g} (e_1) < m \leq 10\,000 \times 1 \text{ g} (e_1)$ $2\,000 \text{ g} < m \leq 10\,000 \text{ g}$




Applied e_2 (2 g)

MPE (g)	Class 3 For loads, m expressed in verification scale intervals, e	MPE (g)	Class 3 (mass)
$\pm 0.5 e_2$	$0 \leq m \leq 500$	$\pm 0.5 \times 2 \text{ g} = \pm 1.0 \text{ g}$	$0 \leq m \leq 500 \times 2 \text{ g} (e_2)$ $0 \text{ g} \leq m \leq 1\,000 \text{ g}$
$\pm 1.0 e_2$	$500 < m \leq 2\,000$	$\pm 1.0 \times 2 \text{ g} = \pm 2.0 \text{ g}$	$500 \times 2 \text{ g} (e_2) < m \leq 2\,000 \times 2 \text{ g} (e_2)$ $1\,000 \text{ g} < m \leq 4\,000 \text{ g}$
$\pm 1.5 e_2$	$2\,000 < m < 10\,000$	$\pm 1.5 \times 2 \text{ g} = \pm 3.0 \text{ g}$	$2\,000 \times 2 \text{ g} (e_2) < m \leq 10\,000 \times 2 \text{ g} (e_2)$ $4\,000 \text{ g} < m \leq 20\,000 \text{ g}$

MPE graph of e_1



MPE graph of e_2



Weight (kg)	MPE (g)
0.5 - 1.0	±1.0
1.0 - 2.0	±1.5
2.0 - 3.0	±2.0
3.0 - 4.0	±2.0
4.0 - 6.0	±3.0

INTERNAL MATERIALS OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIST)

Answer
Specification of NAWI is
Max : 6 kg, e_1 : 1 g (0 – 3 kg) e_2 : 2 g (3 – 6 kg)
*** Multi interval instrument**

MPE (g)	Class 3 (Gram)	Mpe change points
± 0.5 g	$0 \text{ g} \leq m \leq 500 \text{ g}$	500 g
± 1.0 g	$500 \text{ g} < m \leq 2 \text{ kg}$	2 kg
± 1.5 g	$2 \text{ kg} < m \leq 3 \text{ kg}$	3 kg
± 2.0 g	$3 \text{ kg} < m \leq 4 \text{ kg}$	4 kg
± 3.0 g	$4 \text{ kg} < m \leq 6 \text{ kg}$	

INTERNAL MATERIALS OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIST)

Exercise 5

Specification of NAWI is
Max : 150 kg
e : 100 g
Class 3

Exercise 6

Specification of NAWI is
Max : 60 kg
e : 20 g
Class 3

INTERNAL MATERIALS OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIST)

Exercise 7

Specification of NAWI is
Max : 500 g
e : 5 g
Class 4

Exercise 8

Specification of NAWI is
Max : 3 kg
e : 0.1 g
Class 2

INTERNAL MATERIALS OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIST)

Exercise 9

Specification of NAWI is

Max : 1 kg

e : 0.2 g

class 3

Exercise 10

Specification of NAWI is

Max : 15 kg

e₁ : 2 g (0 ~ 6 kg)

e₂ : 5 g (6 ~ 15 kg)

*** Multi interval instrument**

Verification Standards (3.7)

Weights 3.7.1

The standard weights or standard masses used for the verification of an instrument shall not have an error greater than 1/3 of the maximum permissible error of the instrument for the applied.

Max : 3 kg, Verification scale interval (e) : 1 g, Min : 20 g, Class 3

Test points	Test load	MPE (e)	MPE(g)	MPE/3	Weights
Zero	0	0.5 e	0.5 g	0.17 g	-
Min	20 g	0.5 e	0.5 g	0.17 g	M2 (3mg)
500 e	500 g	0.5 e	0.5 g	0.17 g	M2 (75mg)
Any	1 000 g	1 e	1.0 g	0.33 g	M2 (150mg)
2 000 e	2 000 g	1 e	1.0 g	0.33 g	M2 (300mg)
Max	3 000 g	1.5 e	1.5 g	0.5 g	M2 (450mg)

OIML R 111 Weights of classes E1, E2, F1, F2, M1, M1-2, M2, M2-3 and M3

Minimum accuracy class of weights

The accuracy class for weights used as standards for the verification of weights or weighing instruments should be in accordance with the requirements of the relevant OIML Recommendations.

The OIML weight classes are defined as follows:

Class E1: Weights intended to ensure traceability between national mass standards (with values derived from the International Prototype of the kilogram) and weights of class E2 and lower. Class E1 weights or weight sets shall be accompanied by a calibration certificate (see 15.2.2.1).

Class E2: Weights intended for use in the verification or calibration of class F1 weights and for use with weighing instruments of special accuracy class I. Class E2 weights or weight sets shall be accompanied by a calibration certificate (see 15.2.2.2). They may be used as class E1 weights if they comply with the requirements for surface roughness, magnetic susceptibility and magnetization for class E1 weights, and if their calibration certificate gives the appropriate data as specified in 15.2.2.1.

Class F1: Weights intended for use in the verification or calibration of class F2 weights and for use with weighing instruments of special accuracy class I and high accuracy class II.

Class F2: Weights intended for use in the verification or calibration of class M1 and possibly class M2 weights. Also intended for use in important commercial transactions (e.g. precious metals and stones) on weighing instruments of high accuracy class II.


Class M1: Weights intended for use in the verification or calibration of class M2 weights, and for use with weighing instruments of medium accuracy class III.

Class M2: Weights intended for use in the verification or calibration of class M3 weights and for use in general commercial transactions and with weighing instruments of medium accuracy class III.

Class M3: Weights intended for use with weighing instruments of medium accuracy class III and ordinary accuracy class IV.

Classes M1-2 and M2-3 : Weights from 50 kg to 5 000 kg of lower accuracy intended for use with weighing instruments of medium accuracy class III.

Note: The error in a weight used for the verification of a weighing instrument shall not exceed 1/3 of the minimum permissible error for an instrument. These values are listed in section 3.7.1 of OIML R 76 Non-automatic Weighing Instruments (1992).

 Test set up  Inspection environment Set up to solid table Vibration Influence by the wind Preparations for ; Weights Report sheet www.aist.ac.jp / ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIST)	 Test producer Part 1: Self indicating instrument (analog indication) Spring type Part 2: Non-self indicating instrument Lever type www.aist.ac.jp / ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIST)
 Test producer - part 1 Self indicating instrument (analog indication)  Spring type www.aist.ac.jp / ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIST)	 Value of maximum permissible error [Weighing performance test] Test loads (A4.4.1 and 8.3.3) Apply test loads from zero up to and including Max, and similarly remove the test loads back to zero. Determining the verification for weighing tests at least 5 shall be selected. The test loads selected shall include Max, and Min, and values at or near those at which the mpe changes. www.aist.ac.jp / ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIST)


Weighing Performance Test

Determine the test loads and mpe

Example : Max : 3 kg, e : 10 g, Min : 100 g, class 4

Test load		mpe
zero		
Min		
Mpe change 1		
Mpe change 2		
Max		
*Any load		

mpe	Class 4 For loads, m expressed in verification scale intervals, e	Class 4 (Gram)
$\pm 0.5 e$	$0 \leq m \leq 50$	$0 g \leq m \leq 500 g$
$\pm 1.0 e$	$50 < m \leq 200$	$500 g < m \leq 2\ 000 g$
$\pm 1.5 e$	$200 < m < 1\ 000$	$2\ 000 g < m < 10\ 000 g$



Weighing Performance Test

Determine the test loads and mpe

Example : Max : 3 kg, e : 10 g, Min : 100 g, classes 4

Test load		mpe
zero	0 g	???
Min	100 g	???
Mpe change 1	???	???
Mpe change 2	???	???
Max	3 kg	???
Any load	???	???

Weighing Performance Test

Determine the test loads and mpe

Example : Max : 3 kg, e : 10 g, Min : 100 g, class 4

Test load		mpe
zero	0 g	5 g
Min	100 g	5 g
Mpe change 1	500 g	5 g
Mpe change 2	2 kg	10 g
Max	3 kg	15 g
Any load	???	???

Weighing Performance Test

Determine the test loads and mpe

Example : Max : 3 kg, e : 10 g, Min : 100 g, class 4

Test load		mpe
zero	0 g	5 g
Min	100 g	5 g
Mpe change 1	500 g	5 g
Any load	1 kg	10 g
Mpe change 2	2 kg	10 g
Max	3 kg	15 g

Weighing Performance Test


Test loads and MPE are recorded in the test report.

Example : Max : 3 kg, e : 10 g, Min : 100 g, class 4


Load	Indication	Error	Corrected error	Mpe
0				5 g
100 g				5 g
500 g				5 g
1 kg				10 g
2 kg				10 g
3 kg				15 g

Initial zero error

Weighing Performance Test Self indication instruments



Weighing Performance Test



Record these the indication value of the instrument when put on the test load

Load	Indication	Error	Corrected error	Mpe
0	0	1		5 g
100 g	101	101		5 g
500 g	503	503		5 g
1 kg	1 005	1 003		10 g
2 kg	2 005	2 010		10 g
3 kg	3 010			15 g

Initial zero error

Weighing Performance Test

Calculate of Error

$$\text{Error (E)} = I - L$$

I = Indication L = Load

$$\text{Corrected error (Ec)} = E - E_0$$

E = Error of each test load

E₀ = initial zero error

* If initial zero error is not zero, this formula shall be applied.

Record these the indication value of the instrument when put on the test load

Load	Indication	Error	Corrected error	Mpe
0	0	1		5 g
100 g	101	101		5 g
500 g	503	503		5 g
1 kg	1 005	1 003		10 g
2 kg	2 005	2 010		10 g
3 kg	3 010			15 g

Initial zero error

Pass or Fail

Record these the indication value of the instrument when put on the test load

Load	Indication	Error	Corrected error	Mpe
0	0	1	0	1
100 g	101	1	1	1
500 g	503	3	3	3
1 kg	1 005	5	5	3
2 kg	2 005	5	10	5
3 kg	3 010	10	10	15 g

Initial zero error

Pass

Example

Corrected error shall be applied in the case of the following results

Load	Indication		Error	Corrected error	mpe
0	5	5			5 g
100 g	105	105			5 g
500 g	505	505			5 g
1 kg	1 010	1 010			10 g
2 kg	2 015	2 015			10 g
3 kg	3 015				15 g

Pass or Fail

Initial zero error

INTERNATIONAL BUREAU OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (IAIST)

Example

Corrected error shall be applied in the case of the following results

Load	Indication		Error	Corrected error	mpe
0	5	5	5	0	5 g
100 g	105	105	5	0	5 g
500 g	505	505	5	0	5 g
1 kg	1 010	1 010	10	5	10 g
2 kg	2 015	2 015	15	10	10 g
3 kg	3 015		15	10	15 g

Pass

Initial zero error

INTERNATIONAL BUREAU OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (IAIST)

Repeatability (A.4.10)

The difference between the results of several weighings of the same load shall not be greater than the absolute value of the maximum permissible error of the instruments.

Two series of weighings shall be performed, one with a load of about 50% and one with a load close to 100% of Max. Normally no more than 3 weighings on class 3 and 4 or 6 weighings on classes 1 and 2 are necessary.

INTERNATIONAL BUREAU OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (IAIST)

Repeatability test

Self indication instruments : analogue indication

Example : Max : 3 kg, e : 10 g, Min : 100 g, class 4

➤ Determining test load and mpe:

1/2 Max = 1.5 kg, mpe is g
Max = 3 kg, mpe is g

➤ Record in test report

No.	Load	Indication	Load	Indication
1	1.5 kg		3 kg	
2	1.5 kg		3 kg	
3	1.5 kg		3 kg	
	Max-Min		Max-Min	

INTERNATIONAL BUREAU OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (IAIST)

Repeatability test

Self indication instruments : analogue indication

Example : Max : 3 kg, e : 10 g, Min : 100 g, class 4

> Determining test load and mpe:

1/2 Max = 1.5 kg, mpe is 10 g


Max = 3 kg, mpe is 15 g

> Record in test report

No.	Load	Indication	Load	Indication
1	1.5 kg		3 kg	
2	1.5 kg		3 kg	
3	1.5 kg		3 kg	
	Max-Min		Max-Min	

INTERNAL MECHANISM OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIST)

Repeatability test



INTERNAL MECHANISM OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIST)

Repeatability test

No.	Load	Indication	Load	Indication
1	1.5 kg	1500 g	3 kg	2990
2	1.5 kg	1505 g	3 kg	2990
3	1.5 kg	1500 g	3 kg	3000
	Max-Min		Max-Min	

Pass or Fail

INTERNAL MECHANISM OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIST)

Repeatability test

No.	Load	Indication	Load	Indication
1	1.5 kg	1500 g	3 kg	2990
2	1.5 kg	1505 g	3 kg	2990
3	1.5 kg	1500 g	3 kg	3000
	Max-Min	5 g	Max-Min	10 g

Pass

INTERNAL MECHANISM OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIST)

Discrimination test (A3.8)

Self indication instruments : analogue indication

An extra load equivalent to the absolute value of the maximum permissible error for the applied load when gently placed on or withdrawn from the instrument at equilibrium shall cause a permanent displacement of the indicating element corresponding to not less than 0.7 times extra load.

This test shall be performed with three different loads.

Min, 1/2Max, Max

Discrimination test

Self indication instruments : analogue indication

Example : **Max : 3 kg, e : 10 g, Min : 100 g, class 4**

> Determine test load and extra load :

Min = 100 g extra load: g
1/2Max = 1.5 kg extra load: g
Max = 3 kg extra load: g

> Record in test report

Load	Indication I1	Extra load = mpe	Indication I2	I2-I1
100 g				
1.5 kg				
3 kg				

Discrimination test

Self indication instruments : analogue indication

Example : **Max : 3 kg, e : 10 g, Min : 100 g, class 4**

> Determine test load and extra load :

Min = 100 g ; extra load: 5 g
1/2Max = 1.5 kg ; extra load: 10 g
Max = 3 kg ; extra load: 15 g


> Record in test report

Load	Indication I1	Extra load = mpe	Indication I2	I2-I1
100 g		5 g		
1.5 kg		10 g		
3 kg		15 g		

Discrimination test

Self indication instruments : analogue indication

Example : **Max : 3 kg, e : 10 g, Min : 100 g, class 4**



AIIST

Discrimination test

Self indication instruments : analogue indication

Example : *Max* : 3 kg, *e* : 10 g, *Min* : 100 g, *class* 4

VERBAAL INSTITUUT OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIIST)

AIIST

Discrimination test

Self indication instruments : analogue indication

Example : *Max* : 3 kg, *e* : 10 g, *Min* : 100 g, *class* 4

Load	Indication I1	Extra load = mpe	Indication I2	I2-I1
100 g	100 g	5 g	105 g	5 g
1.5 kg	1500 g	10 g	1510 g	10 g
3 kg	3000 g	15 g	3012 g	12 g

Check if $I2-I1 \geq 0,7 \text{ mpe}$

Pass or Fail

VERBAAL INSTITUUT OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIIST)

AIIST

Discrimination test

Self indication instruments : analogue indication

Example : *Max* : 3 kg, *e* : 10 g, *Min* : 100 g, *class* 4

Load	Indication I1	Extra load = mpe	Indication I2	I2-I1
100 g	100 g	5 g	105 g	5 g
1.5 kg	1500 g	10 g	1510 g	10 g
3 kg	3000 g	15 g	3012 g	12 g

Check if $I2-I1 \geq 0,7 \text{ mpe}$

Pass

VERBAAL INSTITUUT OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIIST)

AIIST

Eccentricity tests (A4.7)

The Indications for different positions of a load shall meet the maximum permissible errors, when the is tested according to following.


- > Instrument with a load receptor having *n* points of support, with $n \leq 4$
Test load : (Max + Additive tare effect) / 3
- > Instrument with a load receptor having *n* points of support, with $n > 4$
Test load : (Max + Additive tare effect) / (*n*-1)
- > Instrument with a load receptor subject to minimal off-center loading
Test load : 0.1Max + Additive tare effect
- > Instrument used for weighing rolling loads
Test load : usual rolling load, the heaviest and the most concentrated one which may be weighed, but not exceeding 0.8 times the sum of the Max and Additive tare effect

VERBAAL INSTITUUT OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIIST)

Eccentricity test

Self indication instruments :
analogue indication

Example : Max : 3 kg, e : 10 g,
Min : 100 g, class 4



> Determine type of load receptor , test load and mpe
This instrument has 2 points of support


n points of support with $n \leq 4$

Test load = $(\text{Max} + \text{Additive tare effect}) / 3 = \boxed{?} \text{ kg}$
mpe = $\boxed{?} \text{ g}$

Eccentricity test

Self indication instruments :
analogue indication

Example : Max : 3 kg, e : 10 g,
Min : 100 g, class 4



> Determine type of load receptor , test load and mpe
This instrument has 2 points of support


n points of support with $n \leq 4$

Test load = $(\text{Max} + \text{Additive tare effect}) / 3 = 3 \text{ kg} / 3 = 1 \text{ kg}$
mpe = 10 g

Eccentricity test

> Determine of different position

Surface of the loading receptor is evenly
divided into four




○ Load positions

Eccentricity test

> Record in test report


Location	L	I	E	Ec	mpe
1	0				
1	1 kg				10 g
2	0				
2	1 kg				10 g
3	0				
3	1 kg				10 g
4	0				
4	1 kg				10 g



Eccentricity test

Self indication instruments : analogue indication

Example : Max : 3 kg, e : 10 g, Min : 100 g, class 4



Eccentricity test

Self indication instruments : analogue indication

Spring type: Example : Max : 3 kg, e : 10 g, Min : 100 g, class 4

Location	L	I	E	Ec	mpe
1	0	0 g			
1	1 kg	1000 g			10 g
2	0	0 g			
2	1 kg	1005 g			10 g
3	0	0 g			
3	1 kg	1005 g			10 g
4	0	5 g			
4	1 kg	1010 g			10 g

Pass or Fail

Eccentricity test

Self indication instruments : analogue indication

Spring type: Example : Max : 3 kg, e : 10 g, Min : 100 g, class 4


Location	L	I	E	Ec	mpe
1	0	0 g	0 g		
1	1 kg	1000 g	0 g	0 g	10 g
2	0	0 g	0 g		
2	1 kg	1005 g	5 g	5 g	10 g
3	0	0 g	0 g		
3	1 kg	1005 g	5 g	5 g	10 g
4	0	5 g	5 g		
4	1 kg	1010 g	10 g	5 g	10 g

Pass

Zero-setting device

- The effect does not alter Max
- The accuracy is 0.25 e or 0.5 d on a auxiliary indicating device
- The range is 4 % of Max for zero-setting

Maximum effect



Zero setting accuracy

After zero setting the effect of zero deviation on the result of the weighing shall not be more than $0.25e$; however, on an instrument with auxiliary indicating devices this effect shall be not more than $0.5d$

Self indicating instruments : analogue indication

Spring type: Example : Max : 3 kg, e : 10 g, Min : 100 g, class 4

> Determine of mpe

e = 10 g


mpe = 0.25 e = ? g

Preload (Max)

Zeroing

Applied loads

Remove to load Reading Indication



Self indicating instruments : analogue indication

Spring type: Example : Max : 3 kg, e : 10 g, Min : 100 g, class 4

> Determine of mpe

e = 10 g


mpe = 0.25 e = 2.5 g

Preload (Max)

Zeroing

Applied loads

Remove to load Reading Indication



Test producer - part 2

Non-self indicating instrument



Value of maximum permissible error [Weighing performance test]

Test loads

Apply test loads from zero up to and including Max, and similarly remove the test loads back to zero.

Determining the verification for weighing tests at least 5 shall be selected. The test loads selected shall include Max, and Min, and values at or near those at which the mpe changes.

Weighing Performance Test Determine the test loads and mpe

Example : Max : 150 kg, e : 50 g, Min : 1 kg, class 3


Test load		mpe
zero		
Min		
Mpe change 1		
Mpe change 2		
Max		
*Any load		

Weighing Performance Test Determine the test loads and mpe

Example : Max : 150 kg, e : 50 g, Min : 1 kg, class 3

Test load		mpe
zero	0	25 g
Min	1 kg	25 g
Mpe change 1	25 kg	25 g
Mpe change 2	100 kg	50 g
Max	150 kg	75 g
*Any load	50 kg	50 g

Weighing Performance Test Non self indication instruments Platform instrument



Weighing Performance Test

Test loads and MPE are recorded in the test report.

Example : Max : 150 kg, e : 50 g, Min : 1 kg, classes 3


Load	Indication	Small weights	Error	Mpe
0				25 g
1 kg				25 g
25 kg				25 g
50 kg				50 g
100 kg				50 g
150 kg				75 g

Weighing Performance Test

Zeroing

1. Small Weights put on the load receptor of no-load after preload.

2. Zero setting device to equilibrium position.



Weighing Performance Test

Test loads and MPE are recorded in the test report.

Example : Max : 150 kg, e : 50 g, Min : 1 kg, classes 3

Load	Indication	Small weights	Error	Mpe
0	0		0	25 g
1 kg				25 g
25 kg				25 g
50 kg				50 g
100 kg				50 g
150 kg				75 g

Weighing Performance Test

Test load (Min)

1. Min weight put on load receptor.

2. Confirm to position of indicator.

◆ If indicator is an accord with equilibrium position, the error is 0.

◆ If indicator balanced above equilibrium position, Small weights shall be removed until equilibrium position comes to the indicator. (-)

◆ If indicator balanced below equilibrium position, Small weights shall be put on until equilibrium position comes to the indicator. (+)

Weighing Performance Test

Test loads and MPE are recorded in the test report.

Example : Max : 150 kg, e : 50 g, Min : 1 kg, classes 3

Load	Indication		Small weights		Error	Mpe
0	0	0	0	-10 g		25 g
1 kg	1 000	1 000	+15 g	-10 g		25 g
25 kg	25 000	25 000	+15 g	-10 g		25 g
50 kg	50 000	50 000	+5 g	-10 g		50 g
100 kg	100 000	100 000	-25 g	-40 g		50 g
150 kg	150 000		-50 g			75 g

Weighing Performance Test

Calculate of Error

Non self indicating instruments

$$\text{Error (E)} = I - L + SI$$

I = Indication L = Load SI= small load

Weighing Performance Test

Test loads and MPE are recorded in the test report.

Example : Max : 150 kg, e : 50 g, Min : 1 kg, classes 3

Load	Indication		Small weights		Error	Mpe
0	0	0	0	-10 g		25 g
1 kg	1000	1000	+15 g	-10 g		25 g
25 kg	25000	25000	+15 g	-10 g		25 g
50 kg	50000	50000	+5 g	-10 g		50 g
100 kg	100000	100000	-25 g	-40 g		50 g
150 kg	150000		-50 g			75 g

Pass or Fail

Weighing Performance Test

Test loads and MPE are recorded in the test report.

Example : Max : 150 kg, e : 50 g, Min : 1 kg, classes 3


Load	Indication		Small weights		Error	Mpe
0	0	0	0	-10 g	0	-10 g
1 kg	1000	1000	+15 g	-10 g	15 g	-10 g
25 kg	25000	25000	+15 g	-10 g	15 g	-10 g
50 kg	50000	50000	+5 g	-10 g	5 g	-10 g
100 kg	100000	100000	-25 g	-40 g	-25 g	-40 g
150 kg	150000		-50 g		-50 g	75 g

Pass

Weighing Performance Test Simple method

Evaluation method; error cannot be determined
Non self indicating instruments


- ◆ If indicator balanced above equilibrium position,



INTERNAL SECURITY OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIIST)

Weighing Performance Test Simple method

- ◆ If indicator balanced below equilibrium position,



INTERNAL SECURITY OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIIST)

Repeatability

The difference between the results several weighings of the same load shall not be greater than the absolute value of the maximum permissible error of the instruments.

Two series of weighings shall be performed, one with a load of about 50% and one with a load close to 100% of Max. Normally no more than 3 weighings on class 3 and 4 or 6 weighings on classes 1 and 2 are necessary.

INTERNAL SECURITY OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIIST)

Repeatability Test

Non self indication instruments

Example : Max : 150 kg, e : 50 g, Min : 1 kg, class 3

➤ Determining test load and mpe:

$$\begin{aligned} 1/2 \text{ Max} &= 75 \text{ kg, mpe : ? g} \\ \text{Max} &= 150 \text{ kg, mpe : ? g} \end{aligned}$$

➤ Record in test report

No	Load	Indication	Small weights	Load	Indication	Small weights
1	75 kg			150 kg		
2	75 kg			150 kg		
3	75 kg			150 kg		
	Max-Min			Max-Min		

INTERNAL SECURITY OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIIST)

Repeatability Test

Non self indication instruments

Example : **Max : 150 kg, e : 50 g, Min : 1 kg, class 3**

➤ Determining test load and mpe:


1/2 Max : 75 kg, mpe : 50 g

Max : 150 kg, mpe : 75 g

➤ Record in test report

No	Load	Indication	Small weights	Load	Indication	Small weights
1	75 kg			150 kg		
2	75 kg			150 kg		
3	75 kg			150 kg		
	Max-Min			Max-Min		

Repeatability test



Repeatability test

No	Load	Indication	Small weights	Load	Indication	Small weights
1	75 kg	75.000 kg	+10 g	150 kg	150.000 kg	-50 g
2	75 kg	75.000 kg	+ 50 g	150 kg	150.000 kg	-50 g
3	75 kg	75.000 kg	+ 50 g	150 kg	150.000 kg	-40 g
	Max-Min			Max-Min		

Pass or Fail

Repeatability test

No	Load	Indication	Small weights	Load	Indication	Small weights
1	75 kg	75.000 kg	+10 g	150 kg	150.000 kg	-50 g
2	75 kg	75.000 kg	+ 50 g	150 kg	150.000 kg	-50 g
3	75 kg	75.000 kg	+ 50 g	150 kg	150.000 kg	-40 g
	Max-Min		40 g	Max-Min		10 g

Pass

Discrimination test

Non self indicating instruments

An extra load equivalent to 0.4 times the absolute value of the maximum permissible error for the applied load when gently placed on or withdrawn from the instrument at equilibrium shall produce a visible cause a permanent displacement of the indicating element movement.

This test shall be performed with three different loads.

Min, 1/2Max, Max

Discrimination test

Non self indication instruments

Example : **Max : 150 kg, e : 50 g, Min : 1 kg, class 3**

> Determine test load and extra load :

Min = 1 kg : Extra load = ? g
1/2Max = 75 kg : Extra load = ? g
Max = 150 kg : Extra load = ? g

> Record in test report

Load	Indication I	Extra load =0.4 mpe	Movement
1 kg			
75 kg			
150 kg			

Discrimination test

Non self indication instruments

Example : **Max : 150 kg, e : 50 g, Min : 1 kg, class 3**

> Determine test load and extra load :

Min = 1 kg : Extra load $25\text{ g} \times 0.4 = 10\text{ g}$
1/2Max = 75 kg : Extra load $50\text{ g} \times 0.4 = 20\text{ g}$
Max = 150 kg : Extra load $75\text{ g} \times 0.4 = 30\text{ g}$

> Record in test report

Load	Indication I	Extra load =0.4 mpe	Movement
1 kg		10 g	
75 kg		20 g	
150 kg		30 g	

Discrimination test

Non self indication instruments



AIIST

Discrimination test

Non self indication instruments

Applied to test load

Indicator

Applied to extra load

Indicator

Movement

INTERNAL SECURITY OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIIST)

AIIST

Discrimination test

Non self indication instruments

Example : Max : 150 kg, e : 50 g, Min : 1 kg, class 3

Load	Indication I	Extra load =0.4 mpe	Movement (*)
1 kg	1.000 kg	10 g	+
75 kg	75.000 kg	20 g	+
150 kg	150.000 kg	30 g	+

(*) Mark visible movement by "+"

Pass or Fail

INTERNAL SECURITY OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIIST)

AIIST

Discrimination test

Non self indication instruments

Example : Max : 150 kg, e : 50 g, Min : 1 kg, class 3

Load	Indication I	Extra load =0.4 mpe	Movement (*)
1 kg	1.000 kg	10 g	+
75 kg	75.000 kg	20 g	+
150 kg	150.000 kg	30 g	+

(*) Mark visible movement by "+"

Pass

INTERNAL SECURITY OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIIST)

AIIST

Eccentricity

The Indications for different positions of a load shall meet the maximum permissible errors, when the is tested according to following.

- > Instrument with a load receptor having n points of support, with $n \leq 4$

Test load : (Max + Additive tare effect) / 3
- > Instrument with a load receptor having n points of support, with $n > 4$

Test load : (Max + Additive tare effect) / (n-1)
- > Instrument with a load receptor subject to minimal off-center loading

Test load : 0.1Max + Additive tare effect
- > Instrument used for weighing rolling loads

Test load : usual rolling load, the heaviest and the most concentrated one which may be weighed, but not exceeding 0.8 times the sum of the Max and Additive tare effect

INTERNAL SECURITY OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIIST)

Eccentricity

Non self indication instruments :

Example : Max : 150 kg, e : 50 g,
Min : 1 kg, class 3

> Determine type of load receptor , test load and mpe


This Instrument has 4 points of support

n points of support with $n \leq 4$

Test load = (Max + Additive tare effect) / 3
= kg
mpe = g




● support



Eccentricity

> Record in test report

Location	L	I	SI	E	mpe
1	0				
1	50 kg				50 g
2	0				
2	50 kg				50 g
3	0				
3	50 kg				50 g
4	0				
4	50 kg				50 g



Eccentricity

Non self indication instruments :

Example : Max : 150 kg, e : 50 g,
Min : 1 kg, class 3

> Determine type of load receptor , test load and mpe


This Instrument has 4 points of support

n points of support with $n \leq 4$

Test load = (Max + Additive tare effect) / 3
= 150 kg / 3 = 50 kg
mpe = 50 g



● support



Eccentricity

Non self indication instruments

Example : Max : 150 kg, e : 50 g, Min : 1 kg, class 3



Eccentricity

Non self indication instruments :

Location	L	I	SI	E	mpe
1	0	0	0		
1	50 kg	50.000 kg	0		50 g
2	0	0	0		
2	50 kg	50.000 kg	+5 g		50 g
3	0	0	0		
3	50 kg	50.000 kg	+10 g		50 g
4	0	0	0		
4	50 kg	50.000 kg	-10 g		50 g

Pass or Fail

Eccentricity

Non self indication instruments :

Location	L	I	SI	E	mpe
1	0	0	0		
1	50 kg	50.000 kg	0	0	50 g
2	0	0	0		
2	50 kg	50.000 kg	+5 g	5 g	50 g
3	0	0	0		
3	50 kg	50.000 kg	+10 g	10 g	50 g
4	0	0	0		
4	50 kg	50.000 kg	-10 g	-10 g	50 g

Pass

Sensitivity

Only applied on non self indicating instrument

An extra load equivalent to the absolute value of the MPE for the applied load, shall be placed on the instrument at equilibrium and shall cause a permanent displacement of the indicating element of at least:

- > 1 mm for an instrument of class 1 or 2;
- > 2 mm for an instrument of class 3 or 4 with Max ≤ 30 kg;
- > 5mm for an instrument of class 3 or 4 with Max > 30 kg.

The test shall be performed with a minimum of two difference loads (e.g. zero and Max)

Sensitivity

Non self indication instruments

Example : **Max : 150 kg, e : 50 g, Min : 1 kg, class 3**

> Determine test load and extra load :

Zero = 0 kg : Extra load ? g

Max = 150 kg : Extra load ? g

> Record in test report

Load	Extra load mpe	Displacement of indicating element	Requirement
0			
150 kg			

Sensitivity

Non self indication instruments

Example : *Max : 150 kg, e : 50 g, Min : 1 kg, class 3*

➤ Determine test load and extra load :

Zero = 0 kg : Extra load 25 g

Max = 150 kg : Extra load 75 g

➤ Record in test report

Load	Extra load mpe	Displacement of indicating element	Requirement
0	25 g		5 mm
150 kg	75 g		5 mm

MANUAL INSTRUCTIONS FOR ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIST)

Sensitivity

Non self indication instruments




MANUAL INSTRUCTIONS FOR ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIST)

Sensitivity


Non self indication instruments

Applied to test load



Indicator

Applied to extra load



Indicator
Distance between the
equilibrium positions shall be
measured before and after
loading of extra load

MANUAL INSTRUCTIONS FOR ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIST)

Sensitivity



Non self indication instruments

Example : *Max : 150 kg, e : 50 g, Min : 1 kg, class 3*

Load	Extra load mpe	Displacement of indicating element	Requirement
0	25 g	10 mm	5 mm
150 kg	75 g	9 mm	5 mm

Pass

MANUAL INSTRUCTIONS FOR ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIST)

 Zero setting accuracy After zero setting the effect of zero deviation on the result of the weighing shall not be more than 0.25e; however, on an instrument with auxiliary indicating devices this effect shall be not more than 0.5d INTERNATIONAL BUREAU OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIIST)	 Zero setting accuracy  INTERNATIONAL BUREAU OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIIST)
 Zero setting accuracy Zeroing 1. Small Weights put on the load receptor of no-load after preload. 2. Zero setting device to equilibrium position.  INTERNATIONAL BUREAU OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIIST)	 Zero setting accuracy After the load is remove, Zero indicator shall be confirmed. Confirm to position of indicator. ◆ If indicator is an accord with equilibrium position, the error is 0. ◆ If indicator balanced above equilibrium position, Small weights shall be removed until equilibrium position comes to the indicator. (-) ◆ If indicator balanced below equilibrium position, Small weights shall be put on until equilibrium position comes to the indicator. (+) INTERNATIONAL BUREAU OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIIST)



Zero setting accuracy

Example : Max : 150 kg, e : 50 g, Min : 1 kg, classes 3

zeroing	load	Indication	SI	Error	Mpe
0	150 kg	0	-10 g		12.5g

Pass or Fail

INTERNATIONAL INSTITUTE OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIIST)



Zero setting accuracy

Example : Max : 150 kg, e : 50 g, Min : 1 kg, classes 3

zeroing	load	Indication	SI	Error	Mpe
0	150 kg	0	-10 g	-10 g	12.5g

Pass

INTERNATIONAL INSTITUTE OF ADVANCED INDUSTRIAL SCIENCE AND TECHNOLOGY (AIIST)